



Product Service

Attestation of Compliance

No. E8N 09 08 62545 005

Holder of Certificate: **DA TECH CO., Ltd**
 #710-C, Woolim Lionsvalley, 425
 Cheongcheon-dong, Bupyeong-gu, Incheon 403-858
 REPUBLIC OF KOREA

Name of Object: Electrical equ. for measurement, control and laboratory use
 Digital Breathalyzer Alcoscent

Model(s): DA-8100


Description of Object: Rated input voltage: 3 V d.c. (1.5 V Battery * 2EA)

Tested according to: EN 61326-2-3:2006 Class B

This Attestation of Compliance is issued according to the Directive 2004/108/EC relating to electromagnetic compatibility on a voluntary basis. It confirms that the listed apparatus complies with all essential requirements of the EMC directive and applies only to the sample and its technical documentation submitted to TÜV SÜD Product Service GmbH for testing and certification. See also notes overleaf.

Test report no.: ESTCE-0908-003

Date, 2009-08-13



CE After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.